Notice of Allowability	Application No.	Applicant(s)
	10/001,937	RALPH ET AL.
	Examiner	Art Unit
	Anu Ramana	3733
The MAILING DATE of this communication appeal All claims being allowable, PROSECUTION ON THE MERITS IS herewith (or previously mailed), a Notice of Allowance (PTOL-85) NOTICE OF ALLOWABILITY IS NOT A GRANT OF PATENT RI	(OR REMAINS) CLOSED in this apply or other appropriate communication IGHTS. This application is subject to	plication. If not included will be mailed in due course. THIS
1. This communication is responsive to the amendment filed	<u>on 5/22/06</u> .	
2. The allowed claim(s) is/are 10-13 and 17-22.		
 3. Acknowledgment is made of a claim for foreign priority urestimated as a claim for foreign priority urestimated as a claim for foreign priority urestimated. a) All b) Some* c) None of the: 1. Certified copies of the priority documents have 2. Certified copies of the priority documents have 3. Copies of the certified copies of the priority do International Bureau (PCT Rule 17.2(a)). * Certified copies not received: Applicant has THREE MONTHS FROM THE "MAILING DATE" noted below. Failure to timely comply will result in ABANDONM THIS THREE-MONTH PERIOD IS NOT EXTENDABLE. 4. A SUBSTITUTE OATH OR DECLARATION must be submained in the submained part of the priority documents and the submained part of the priority documents and the submained part of the priority documents and the submained part of the priority documents have a claim of the p	e been received. e been received in Application No cuments have been received in this of this communication to file a reply MENT of this application. Initted. Note the attached EXAMINER es reason(s) why the oath or declarate st be submitted. Is son's Patent Drawing Review (PTO- Is Amendment / Comment or in the Coll. Is Amendment / Comment or in the Coll. Is Af(c)) should be written on the drawing the header according to 37 CFR 1.121(complying with the requirements 'S AMENDMENT or NOTICE OF tion is deficient. 948) attached Office action of the back) of d). must be submitted. Note the
Attachment(s) 1. Notice of References Cited (PTO-892) 2. Notice of Draftperson's Patent Drawing Review (PTO-948) 3. Information Disclosure Statements (PTO-1449 or PTO/SB/O Paper No./Mail Date 4. Examiner's Comment Regarding Requirement for Deposit of Biological Material	6. Interview Summary Paper No./Mail Da 7. Examiner's Amenda 8. Examiner's Statem 9. Other	te

Art Unit: 3733

EXAMINER'S AMENDMENT

An examiner's amendment to the record appears below. Should the changes and/or additions be unacceptable to applicant, an amendment may be filed as provided by 37 CFR 1.312. To ensure consideration of such an amendment, it MUST be submitted no later than the payment of the issue fee.

Authorization for this examiner's amendment was given in a telephone interview with the Attorney of Record, Daniel D. Ryan, on August 3, 2006. See attached Interview Summary Form.

The application has been amended as follows:

In claim 17, line 12: insert - - in the side wall - - after "opening" to provide correct antecedent basis and avoid confusion.

The following is an examiner's statement of reasons for allowance.

Regarding claim 10, none of the prior art either alone or in combination, discloses the method steps of: providing a cannula including a side wall, an opening in the side wall of the cannula, the opening extending partially about the side wall and being elongated along the axis of the cannula and a bone engaging structure on the distal region of the cannula spaced distally of the opening; placing the cannula in bone such that the bone engaging structure engages the cortical bone; inserting an expandable structure through the internal bore of the cannula into registration with the opening; and expanding the expandable structure through the opening in the side wall of the cannula into contact with cancellous bone.

Regarding claim 17, none of the prior art either alone or in combination, discloses the method steps of: providing a cannula including a side wall, an opening in the side wall of the cannula, the opening extending partially about the side wall and being elongated along the axis of the cannula and a distal opening in the distal region of the cannula communicating with an internal bore of the cannula; introducing a guide pin into bone; introducing the cannula into the bone by passing the guide pin through the distal opening and the internal bore; withdrawing the guide pin; inserting an expandable structure through the internal bore of the cannula into registration with the opening in

Art Unit: 3733

the side wall; and expanding the expandable structure through the opening in the side wall into contact with cancellous bone.

Regarding claim 21, none of the prior art either alone or in combination, discloses the method steps of: providing a cannula including a side wall, an opening in the side wall of the cannula, the opening extending partially about the side wall and being elongated along the axis of the cannula and an internal bore in the cannula terminating at the distal terminus; inserting an expandable structure through the internal bore of the cannula into registration with the opening in the side wall; expanding the expandable structure through the opening in the side wall into contact with cancellous bone to form a cavity; and flowing a volume of filling material into the cavity.

Any comments considered necessary by applicant must be submitted no later than the payment of the issue fee and, to avoid processing delays, should preferably accompany the issue fee. Such submissions should be clearly labeled "Comments on Statement of Reasons for Allowance."

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Anu Ramana whose telephone number is (571) 272-4718. The examiner can normally be reached Monday through Friday between 8:00 am to 5:00 pm.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Eduardo Robert can be reached at (571) 272-4719. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Application/Control Number: 10/001,937

Art Unit: 3733

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see http://pair-direct.uspto.gov. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

Page 4

AR

August 3, 2006

EDUARDO/C/ROBERT SUPERVISORY PATIÉNT EXAMINER